



Gram Negative Susceptibility Card

SUMMARY AND EXPLANATION

The VITEK® 2 Gram Negative Susceptibility Card is intended for use with the VITEK 2 Systems in clinical laboratories as an in vitro test to determine the susceptibility of **clinically significant aerobic gram-negative bacilli** to antimicrobial agents when used as instructed in the Product Information manual.

INSTRUCTIONS FOR USE

A package insert is provided in the kit or downloadable from www.biomerieux.com/techlib

See the Product Information manual for additional Instructions for Use.

STORAGE CONDITIONS

Store at 2° to 8° C.

CONTENTS OF THE CARD

Antimicrobial	Code	Concentration §	Calling Range		FDA Indications for Use
			≤	≥	
Ampicillin	AM	4, 8, 32	2	32	CSAGNB**
Ampicillin/Sulbactam	SAM	4/2, 16/8, 32/16	2/1	32/16	CSAGNB**
Cefazolin	CZ	4, 16, 64	4	64	CSAGNB**
Cefepime	FEP	0.25, 1, 4, 16, 32	0.12	32	<i>Enterobacter</i> spp., <i>E. coli</i> , <i>K. pneumoniae</i> , <i>P. mirabilis</i> , <i>P. aeruginosa</i> , <i>C. koseri</i> , <i>C. freundii</i> , <i>P. agglomerans</i> , <i>K. oxytoca</i> , <i>P. vulgaris</i> , <i>Pv. rettgeri</i> , <i>Pv. stuartii</i> , <i>S. marcescens</i>
Ceftazidime	CAZ	1, 2, 8, 32	1	64	CSAGNB**
Ceftriaxone	CRO	0.12, 0.25, 1, 4, 16	0.25	64	<i>E. aerogenes</i> , <i>E. coli</i> , <i>K. oxytoca</i> , <i>K. pneumoniae</i> , <i>P. mirabilis</i> , <i>S. marcescens</i> , <i>C. koseri</i> , <i>C. freundii</i> , <i>Shigella</i> spp., PROV(+ <i>Pv. rettgeri</i>)**, SAL(+ <i>S. typhi</i>)**
Ciprofloxacin	CIP	0.5, 2, 4	0.25	4	CSAGNB**
Ertapenem	ETP	0.03, 0.12, 0.5, 2	0.12	8	<i>E. coli</i> , <i>K. pneumoniae</i> , <i>C. freundii</i> , <i>C. koseri</i> , <i>E. aerogenes</i> , <i>E. cloacae</i> , Kox(-ESBL)**, <i>M. morgani</i> , <i>P. mirabilis</i> , <i>P. vulgaris</i> , <i>Pv. rettgeri</i> , <i>Pv. stuartii</i> , <i>S. marcescens</i>
ESBL	ESB	FEP 1, CTX 0.5, CAZ 0.5, FEP/CA 1/10, CTX/CA 0.5/4, CAZ/CA 0.5/4	NEG	POS	<i>E. coli</i> , <i>K. pneumoniae</i> , <i>K. oxytoca</i>
Gentamicin	GM	4, 16, 32	1	16	CSAGNB**
Imipenem	IPM	1, 2, 6, 12	0.25	16	<i>Acinetobacter</i> spp., <i>Citrobacter</i> spp., <i>Enterobacter</i> spp., <i>E. coli</i> , <i>Klebsiella</i> spp., <i>M. morgani</i> , <i>P. vulgaris</i> , <i>Pv. rettgeri</i> , <i>P. aeruginosa</i> , <i>S. marcescens</i> , <i>Pv. stuartii</i>
Levofloxacin	LEV [®]	0.25, 0.5, 2, 8	0.12	8	<i>E. cloacae</i> , <i>E. coli</i> , <i>K. pneumoniae</i> , <i>P. mirabilis</i> , <i>P. aeruginosa</i> , <i>S. marcescens</i> , <i>A. baumannii</i> , <i>A. lwoffii</i> , <i>C. koseri</i> , <i>C. freundii</i> , <i>E. aerogenes</i> , <i>E. sakazakii</i> , <i>K. oxytoca</i> , <i>M. morgani</i> , <i>P. agglomerans</i> , <i>P. vulgaris</i> , <i>Pv. rettgeri</i> , <i>Pv. stuartii</i> , <i>P. fluorescens</i> , <i>C. sakazakii</i>
Nitrofurantoin	FT	16, 32, 64	16	512	CSAGNB**

Antimicrobial	Code	Concentration §	Calling Range		FDA Indications for Use
			≤	≥	
Piperacillin/Tazobactam ^c	TZP	2/4, 8/4, 24/4, 32/4, 32/8, 48/8	4/4	128/4	<i>A. baumannii</i> , <i>E. coli</i> , <i>K. pneumoniae</i> , <i>P. aeruginosa</i> , <i>C. koseri</i> , <i>M. organii</i> , <i>P. mirabilis</i> , <i>P. vulgaris</i> , <i>Pv. rettgeri</i> , <i>Pv. stuartii</i> , <i>S. enterica</i>
Tobramycin	TM	8, 16, 64	1	16	CSAGNB**
Trimethoprim/Sulfamethoxazole	SXT [Ⓢ]	1/19, 4/76, 16/304	20 (1/19)	320 (16/304)	<i>Klebsiella</i> spp., <i>Enterobacter</i> spp., <i>M. organii</i> , <i>P. vulgaris</i> , <i>P. mirabilis</i> , <i>S. sonnei</i> , <i>S. flexneri</i> , Eco(+ETEC)**, <i>C. sakazakii</i>

Numerical values are expressed in µg/ml.

§ Equivalent standard method concentration by efficacy.

**CSAGNB = Clinically significant aerobic gram negative bacilli

**PROV(+*Pv. rettgeri*) = *Providencia* spp. (including *Pv. rettgeri*)

**SAL(+*S. typhi*) = *Salmonella* spp. (including *S. typhi*)

**Kox(-ESBL) = *K. oxytoca* (excluding ESBL producing strains)

Note: For ESBL, FEP is Cefepime, CTX is Cefotaxime, CAZ is Ceftazidime and CA is Clavulanic Acid. For ESBL positive strains, the test interpretation should be reported as resistant for all penicillins, cephalosporins, and aztreonam.

NEG = Negative

POS = Positive

A negative ESBL test result does not rule out the presence of an ESBL masked by an AmpC beta-lactamase.

Ⓢ, Ⓣ etc. = See performance characteristics identified by the drug code with this symbol in the Comment column in Systems Product Information.

^c = Category agreement was established at the time of FDA clearance. Essential agreement was not established since test contains less than five discrete dilutions.

**Eco(+ETEC) = *E. coli* (including susceptible enterotoxigenic strains implicated in traveler's diarrhea)

QUALITY CONTROL

Antimicrobial	Code	CLSI® Quality Control Organisms VITEK 2 Results			
		<i>E. coli</i> ATCC® 25922™	<i>P. aeruginosa</i> ATCC® 27853™	<i>E. coli</i> ATCC® 35218™	<i>K. pneumoniae</i> ssp <i>pneumoniae</i> ATCC® 700603™
Ampicillin	AM	≤2 – 8	-	-	-
Ampicillin/Sulbactam	SAM	≤2/1 – 8/4	-	8/4 – ≥32/16	-
Cefazolin	CZ	≤4	-	-	-
Cefepime	FEP	≤0.12* (*FDA/CLSI Broth Microdilution expected QC range = 0.015 – 0.12 µg/ml)	0.5 – 4	-	-
Ceftazidime	CAZ	≤1	≤1 – 4	-	-
Ceftriaxone	CRO	≤0.25* (*FDA/CLSI Broth Microdilution expected QC range = 0.03 – 0.12 µg/ml)	8 – ≥64	-	-
Ciprofloxacin	CIP	≤0.25	≤0.25 – 1	-	-
Ertapenem	ETP	≤0.12 †	-	-	-
ESBL	ESB	NEG	-	-	POS
Gentamicin	GM	≤1	≤1 – 2	-	-
Imipenem	IPM	≤0.25	1 – 4	-	-
Levofloxacin	LEV	≤0.12	0.5 – 4	-	-
Nitrofurantoin	FT	≤16	-	-	-
Piperacillin/Tazobactam ^c	TZP	≤4/4* (*FDA/CLSI Broth Microdilution expected QC range = 1/4 – 4/4 µg/ml)	≤4/4 – 8/4* (*FDA/CLSI Broth Microdilution expected QC range = 1/4 – 8/4 µg/ml)	≤4/4* (*FDA/CLSI Broth Microdilution expected QC range = 0.5/4 – 2/4 µg/ml)	-
Tobramycin	TM	≤1	≤1	-	-
Trimethoprim/Sulfamethoxazole	SXT	≤20 (1/19)	160 (8/152) – ≥320 (16/304)	-	-

Numerical values are expressed in µg/ml.

† = Does not include the full CLSI/FDA-recommended dilution range for QC testing with this organism.

Note: For ESBL, FEP is Cefepime, CTX is Cefotaxime, CAZ is Ceftazidime and CA is Clavulanic Acid. For ESBL positive strains, the test interpretation should be reported as resistant for all penicillins, cephalosporins, and aztreonam.

NEG = Negative

POS = Positive

A negative ESBL test result does not rule out the presence of an ESBL masked by an AmpC beta-lactamase.

^c = Category agreement was established at the time of FDA clearance. Essential agreement was not established since test contains less than five discrete dilutions.

LIMITATIONS

Perform an alternative method of testing prior to reporting of results for the following antibiotic/organism combination(s):

- Ampicillin: *Citrobacter* spp., *Enterobacter* spp., *Pantoea* spp., *Serratia* spp., *Cronobacter sakazakii*
- Ampicillin/Sulbactam: *Citrobacter* spp., *Enterobacter* spp., *Pantoea* spp., *Serratia* spp., *Cronobacter sakazakii*
- Cefepime: *Bordetella bronchiseptica*, *Hafnia alvei*, *Morganella* spp.
- Ceftriaxone: *Proteus vulgaris*, *Enterobacter cloacae*, *E. cloacae* complex, *Morganella* spp.
- Ertapenem: *Hafnia alvei*
- Imipenem: *Serratia marcescens*
- Piperacillin/Tazobactam: *Serratia marcescens*

Perform an alternative method of testing prior to reporting results when a resistant result is obtained with the following antibiotic/organism combination(s):

- Imipenem: *Aeromonas* spp.
- Piperacillin/Tazobactam: *Pseudomonas aeruginosa*

The ability of the AST card to detect resistance with the following combination(s) is unknown because resistant strains were not available at the time of comparative testing:

- Ceftriaxone: *Shigella* spp., *Providencia rettgeri*, *Salmonella* spp.

Although within Essential Agreement, the lack of an intermediate category has shown major and very major discrepancies when compared to the reference method. Testing should be repeated using an alternative testing/reference method prior to reporting results when the VITEK 2 MIC is 8 µg/ml or 16 µg/ml for the following antibiotic/organism(s):

- Cefepime: *Pseudomonas aeruginosa*

Due to an insufficient number of on-scale isolates available for comparative testing, the performance of VITEK 2 Gram Negative **Ertapenem** is unknown for isolates with MICs in the range of 0.25 - 0.5 µg/mL. To avoid the occurrence of very major errors, isolates with MICs of 0.25 µg/mL and 0.5 µg/mL should be retested using another method.

NOTE: A result for an antibiotic/organism combination which may have a limitation may be suppressed from reporting. Refer to the software user manual for instructions.

↓ MUST enter the following barcodes into "Flex Panel Entry" program before first use of this Susceptibility Card.



INDEX OF SYMBOLS

Symbol	Meaning
	Catalogue number
	In Vitro Diagnostic Medical Device
	Manufacturer
	Date of Manufacture
	Temperature limitation
	Use by
	Batch code
	Consult Instructions for Use
	Contains sufficient for <n> tests
	Authorized representative in the European Community

PATENTS

Product covered by one or more of U.S. Patent Nos. D414,272; D437,797; 5,609,828; 5,746,980; 5,804,437; 5,869,005; 5,932,177; 5,951,952; 6,267,929; 6,309,890 and 6,340,573; and Foreign Counterparts. Other Patents Pending.

WARRANTY

bioMérieux disclaims all warranties, express or implied, including any implied warranties of MERCHANTABILITY AND FITNESS FOR A PARTICULAR USE. bioMérieux shall not be liable for any incidental or consequential damages. IN NO EVENT SHALL BIOMERIEUX'S LIABILITY TO CUSTOMER UNDER ANY CLAIM EXCEED A REFUND OF THE AMOUNT PAID TO BIOMERIEUX FOR THE PRODUCT OR SERVICE WHICH IS THE SUBJECT OF THE CLAIM.

BIOMERIEUX, the blue logo, VITEK and bioLiaison are used, pending and/or registered trademarks belonging to bioMérieux SA or one of its subsidiaries, or one of its companies. The ATCC trademark and trade name and any and all ATCC catalog numbers are trademarks of the American Type Culture Collection. Any other name or trademark is the property of its respective owner.



 **bioMérieux, Inc.**
100 Rodolphe Street
Durham, North Carolina
27712 USA

 **bioMérieux SA**
Chemin de l'Orme
RCS LYON 673 620 399
69280 Marcy-l'Etoile / France
Tel. 33 (0)4 78 87 20 00
Fax 33 (0)4 78 87 20 90
www.biomerieux.com

